



MolyVer® 6 molybdenum reagent AccuVac® ampoules, 25/pk

Nr. produs:

2522025

Produse periculoase

ARTICOL ÎNVECHIT

Acest articol nu mai este disponibil.



Produse înlocuitoare sugerate:

- 2604100

For determination of HR Molybdate Molybdenum by the Mercaptoacetic Acid Method. Hach method 8036. Range: 0.2 - 40.0 mg/L Mo. Pack of 25 ampoules. Also requires CDTA Solution, 0.4 M, Cat. No. 26154-36.

Fast

AccuVac® Ampoules provide quick results. Just snap, fill and read - saving you valuable time.

Easy

No need to measure sample or reagents. The evacuated ampoules draw in the correct sample volume, so anyone can obtain reliable results with minimal effort.

Accurate

Optical quality evacuated vials assure precise measurements and reagent stability.

Use them anywhere you need results.

Ampoules are both reagent containers and photometric measurement cells. No need to transport reagents or samples.

Specificații

Domeniu de măsurare:	0.2 - 40.0 mg/L Mo
Numar teste:	25
Nume metoda:	Mercaptoacetic Acid
Parametru:	Molybdenum
Platforma:	Accuvac